

# General European OMCL Network (GEON) QUALITY MANAGEMENT DOCUMENT

## PA/PH/OMCL (08) 73 R7

### QUALIFICATION OF EQUIPMENT CORE DOCUMENT

<b>Full document title and reference</b>	Qualification of Equipment – Core document <i>PA/PH/OMCL (08) 73 R7</i>
<b>Document type</b>	Guideline
<b>Legislative basis</b>	Council Directive 2001/83/EC and Regulation (EU) 2019/6, as amended
<b>Date of first adoption</b>	1 October 1999
<b>Date of original entry into force</b>	1 February 2000
<b>Date of entry into force of revised document</b>	15 November 2023
<b>Previous titles/other references / last valid version</b>	This document replaces document PA/PH/OMCL (08) 73 R5
<b>Custodian Organisation</b>	The present document was elaborated by the OMCL Network / EDQM of the Council of Europe
<b>Concerned Network</b>	GEON

N.B. This OMCL Quality Management System document is applicable to members of the European OMCL Network only. Other laboratories might use the document on a voluntary basis. However, please note that the EDQM cannot treat any questions related to the application of the documents submitted by laboratories other than the OMCLs of the Network.

# QUALIFICATION OF EQUIPMENT

## CORE DOCUMENT

**Note: Mandatory requirements in this document are defined using the terms “shall” or “must”. The use of “should” indicates a recommendation. For these parts of the text other appropriately justified approaches are acceptable. The term “can” indicates a possibility or an example with non-binding character.**

### Contents

1.	Introduction .....	2
2.	LEVEL I. Selection of instruments and suppliers .....	3
3.	LEVEL II. Installation and release for use .....	3
4.	LEVEL III. Periodic and motivated instrument checks .....	3
5.	LEVEL IV. In-use instrument checks .....	3
6.	List of instrument-related annexes.....	3
7.	LEVEL I. Selection of instruments and suppliers Example of checklist (non-exhaustive)....	4
8.	LEVEL II. Installation and release for use Example of checklist (non-exhaustive) .....	6

### 1. Introduction

The standard ISO/IEC 17025 requires that a laboratory shall have access to equipment that is required for the correct performance of laboratory activities. In particular, calibration programmes shall be established, reviewed and adjusted, including intermediate checks to maintain confidence in the calibration status.

In order to guarantee harmonised interpretation and application of ISO/IEC 17025 requirements within the OMCL Network, the guideline “Qualification of Equipment” has been elaborated.

This document should be considered as a guide to OMCLs for planning, performing and documenting the equipment qualification process. It contains the general introduction and general forms for Level I (Selection of instruments and suppliers) and Level II (Installation and release for use) of qualification, which are common to all types of equipment.

Level III (Periodic and motivated instrument calibration/checks) and Level IV (In-use instrument checks) qualification requirements can be found in separate equipment-related annexes. Level III and IV of qualification must be carried out according to ISO/IEC 17025. Requirements and (if applicable) corresponding typical acceptance limits (given in bold) should be applied; however, other appropriately justified approaches are acceptable. Exemplary procedures provided in the annexes are not binding. They can be helpful when carrying out the required qualification. Nevertheless, other procedures can be applied depending, for example, on the type/model of the equipment.

If the qualification of equipment is done by the manufacturer or an external service provider, it is the responsibility of the OMCL to make sure that this is in line with the requirements set out in this guideline and in the equipment-specific annexes.

The following four levels of Equipment Qualification should be considered by the OMCLs:

## **2. LEVEL I. Selection of instruments and suppliers**

The selection and purchase of new instruments shall follow a documented decision process, based on the needs related to the intended use of the instrument.

An example checklist for setting and documenting such specifications and decisions taken is given in Paragraph 7.

## **3. LEVEL II. Installation and release for use**

When receiving an instrument, the OMCL should check its condition and fulfilment of the order, and monitor the installation process in the selected environment. This includes start-up checks and a full qualification (normally carried out by the supplier) which fulfils (at minimum) the requirements given in Level III.

The release for use shall be authorised by qualified staff, in order to verify that the equipment conforms to the requirements. This decision may be based on the documentation provided by the supplier.

An example checklist for documenting the instrument installation and release for use and decisions taken is given in Paragraph 8.

## **4. LEVEL III. Periodic and motivated instrument checks**

When instruments are installed or moved into a new environment, or after significant repair or maintenance operations, a series of calibrations/checks shall be carried out to maintain confidence in the performance of the equipment, according to a defined procedure. Unless defined in the equipment-specific annexes, the frequency of periodic checks shall be based on pre-defined criteria and documented; where adjustments are made, sound scientific justification shall be provided and documented.

## **5. LEVEL IV. In-use instrument checks**

During the routine use of the instruments, a series of calibrations/checks shall be carried out to maintain confidence in the performance of the equipment and compliance with the system suitability criteria, according to a defined procedure, e.g. specific analytical method, Ph. Eur. chapter/monograph or manufacturer dossier.

In the case of OMCLs performing routine testing (e.g. batch release of vaccines and blood products), the use of control charts can provide supplementary information on equipment performance, which can also be used in this context.

### **NOTE:**

From experience, the terms DQ (Design Qualification), IQ (Installation Qualification), OQ (Operational Qualification) and PQ (Performance Qualification) (not explicitly mentioned by ISO/IEC 17025) have been used in a non-harmonised way by the OMCLs. Therefore, these terms have not been used in this document. This does not exclude their use in OMCL quality systems.

## **6. List of instrument-related annexes**

The up-to-date list of instrument-related annexes, mainly dealing with Level III and Level IV qualification, is available on the EDQM website: <https://www.edqm.eu/en/quality-management-qm-documents>.

**7. LEVEL I. Selection of instruments and suppliers**  
**Example of checklist (non-exhaustive)**

Manufacturer:

Supplier:

Name of instrument and type:

<b>Attribute</b> (This list may be adapted if necessary)	<b>Specifications</b>	<b>Benefits</b> (Instrument/supplier)	<b>Assessment</b>	
			<b>Pass</b>	<b>Fail</b>
<b>Technique</b>				
<b>Communication and data handling</b>				
Interface RS232				
Data transfer to spreadsheets				
Compatible with other hard- and software such as LIMS...				
<b>Safety</b>				
Irradiation				
Explosion protection				
<b>Documentation</b>				
Manual (paper or e-copy)				
<b>Handling</b>				
User language				
<b>Service and maintenance</b>				
Services offered				
Warranty				
<b>Support</b>				
Delivery (duration etc.)				
Installation (Service/Laboratory/Supplier)				
Training (in-house/external courses)				

**Requirements for media and environment**

**Cost/Benefit Analysis**

**Comments/Decisions**

**Date/Signature:**

Date/Signature for approval:

**8. LEVEL II. Installation and release for use**  
**Example of checklist (non-exhaustive)**

Name of instrument and type:

Identification code:

**Conformity with order (instrument/material/documentation)**

Pass	Fail (description of deficiencies)

**Check for damage**

Pass	Fail (description of deficiencies)

**Check of required media supply (connections/environmental conditions)**

Pass	Fail (description of deficiencies)

**Installation of instrument(s) including possible control modules**

Pass	Fail (description of deficiencies)

**Performance of start-up checks and diagnostic functions<sup>1</sup>**

Pass	Fail (description of deficiencies)

<sup>1</sup> if available

**Comments**

--

Where appropriate, raw data are attached to the instrument documentation.

**Date/Signature:**

**Release and authorisation for use:**

**Date/Signature:**

**Disclaimer:**

The present Core Document of the OMCL Guideline “Qualification of Equipment”, as well as its annexes, have been drafted by ad-hoc working groups of technical experts, mainly from Official Medicines Control Laboratories (OMCLs) and only occasionally from other public institutions. These working groups do not include any representative from any commercial organisation.

This Core Document and its Annexes may contain trade names of laboratory instruments, materials and/or reagents. These are exclusively given as examples in order to make these guidelines easier to understand and implement, and were found to be suitable when the guideline was being developed. These references do not in any way imply that the mentioned instruments, materials or reagents, or their suppliers, are especially endorsed, recommended or certified by the EDQM, the OMCL Network or the Council of Europe, in preference to others of a similar nature which are not mentioned. It is therefore acceptable to use instruments, materials and reagents from another source, provided that they fulfil the necessary criteria laid down in these documents and appropriately satisfy the needs of the laboratories concerned in the frame of their specific activities.