

**PHARMACOPOEIAL DISCUSSION GROUP****CORRECTION****CODE: B-01****NAME: Amino Acid Determination****(Correction of the sign-off document Revision 1 signed on 26 October 2016)**

Item to be corrected:

"Methodologies of Amino Acid Analysis General Principles"

The last sentence in "METHOD 8- Precolumn NBD-F Derivatization General Principle" was changed as below;

"Profile analysis ~~was~~ can be achieved for about 1.5 ~~mg~~ µg of protein hydrolysates in the final precolumn labeling reaction mixture for HPLC."**Harmonised provisions:**

<b>Provision</b>	<b>EP</b>	<b>JP</b>	<b>USP</b>
Introduction	+	+	+
Apparatus	+	+	+
General Precautions	+	+	+
Reference Standard Material	+	+	+
Calibration of Instrumentation	+	+	+
Repeatability	+	+	+
Sample Preparation	+	+	+
Internal Standards	+	+	+
Protein Hydrolysis	+	+	+
Methodologies of Amino Acid Analysis General Principles	+	+	+
Data Calculation and Analysis	+	+	+

+ will adopt and implement; - will not stipulate

**Non-harmonised provisions:**

None


**Local requirements**

EP	JP	USP
None	None	Footnote : <i>“Suitable standards may be obtained from NIST (Gaithersburg, MD), Beckman Instruments (Fullerton, CA), Sigma Chemical (St. Louis, MO), Pierce (Rockford, IL), or Agilent (Palo Alto, CA).”</i>  APPENDIX

**European Pharmacopoeia**

Signature	Name	Date
Signé par :  5D202E6E19D1466...	cathie vielle	6 Dec 2024

**Japanese Pharmacopoeia**

Signature	Name	Date
署名者:  for K. Nakaz 9BF72DA462C9442...	Yoshiro Saito	28 Nov. 2024

**United States Pharmacopoeia**

Signature	Name	Date
Signed by:  A7467E52FCC94E9...	Kevin Moore	11/27/2024



33 transforming the analog signal from the detector and for quantitation. It is  
34 preferred that instrumentation be dedicated particularly for amino acid analysis.

35

36

### **General Precautions**

37

38

39 Background contamination is always a concern for the analyst in  
40 performing amino acid analysis. High purity reagents are necessary (e.g., low  
41 purity hydrochloric acid can contribute to glycine contamination). Analytical  
42 reagents are changed routinely every few weeks using only high-pressure liquid  
43 chromatography (HPLC) grade solvents. Potential microbial contamination and  
44 foreign material that might be present in the solvents are reduced by filtering  
45 solvents before use, keeping solvent reservoirs covered, and not placing amino  
46 acid analysis instrumentation in direct sunlight.

46

47

48

49

50

51

52

53

54

55

56

57

58

59

### **Reference Standard Material**

60

61

62

63

64

Acceptable amino acid standards are commercially available for amino  
acid analysis and typically consist of an aqueous mixture of amino acids. When  
determining amino acid composition, protein or peptide standards are analyzed  
with the test material as a control to demonstrate the integrity of the entire

65 procedure. Highly purified bovine serum albumin has been used as a protein  
66 standard for this purpose.

67

## 68 **Calibration of Instrumentation**

69

70 Calibration of amino acid analysis instrumentation typically involves  
71 analyzing the amino acid standard, which consists of a mixture of amino acids at a  
72 number of concentrations, to determine the response factor and range of analysis  
73 for each amino acid. The concentration of each amino acid in the standard is  
74 known. In the calibration procedure, the analyst dilutes the amino acid standard  
75 to several different analyte levels within the expected linear range of the amino  
76 acid analysis technique. Then, replicates at each of the different analyte levels  
77 can be analyzed. Peak areas obtained for each amino acid are plotted versus the  
78 known concentration for each of the amino acids in the standard dilution. These  
79 results will allow the analyst to determine the range of amino acid concentrations  
80 where the peak area of a given amino acid is an approximately linear function of  
81 the amino acid concentration. It is important that the analyst prepare the samples  
82 for amino acid analysis so that they are within the analytical limits (e.g., linear  
83 working range) of the technique employed in order to obtain accurate and  
84 repeatable results.

85

86 Four to six amino acid standard levels are analyzed to determine a  
87 response factor for each amino acid. The response factor is calculated as the  
88 average peak area or peak height per nmol of amino acid present in the standard.  
89 A calibration file consisting of the response factor for each amino acid is prepared  
90 and used to calculate the concentration of each amino acid present in the test  
91 sample. This calculation involves dividing the peak area corresponding to a given  
92 amino acid by the response factor for that amino acid to give the nmol of the  
93 amino acid. For routine analysis, a single-point calibration may be sufficient;  
94 however, the calibration file is updated frequently and tested by the analysis of  
95 analytical controls to ensure its integrity.

95

96

## Repeatability

97

98

Consistent high quality amino acid analysis results from an analytical laboratory require attention to the repeatability of the assay. During analysis of the chromatographic separation of the amino acids or their derivatives, numerous peaks can be observed on the chromatogram that correspond to the amino acids. The large number of peaks makes it necessary to have an amino acid analysis system that can repeatedly identify the peaks based on retention time and integrate the peak areas for quantitation. A typical repeatability evaluation involves preparing a standard amino acid solution and analyzing many replicates (i.e., six analyses or more) of the same standard solution. The relative standard deviation (RSD) is determined for the retention time and integrated peak area of each amino acid. An evaluation of the repeatability is expanded to include multiple assays conducted over several days by different analysts. Multiple assays include the preparation of standard dilutions from starting materials to determine the variation due to sample handling. Often the amino acid composition of a standard protein (e.g., bovine serum albumin) is analyzed as part of the repeatability evaluation. By evaluating the replicate variation (i.e., RSD), the laboratory can establish analytical limits to ensure that the analyses from the laboratory are under control. It is desirable to establish the lowest practical variation limits to ensure the best results. Areas to focus on to lower the variability of the amino acid analysis include sample preparation, high background spectral interference due to quality of reagents and/or laboratory practices, instrument performance and maintenance, data analysis and interpretation, and analyst performance and habits. All parameters involved are fully investigated in the scope of the validation work.

121

122

## Sample Preparation

123

124

125

126

127

128

Accurate results from amino acid analysis require purified protein and peptide samples. Buffer components (e.g., salts, urea, detergents) can interfere with the amino acid analysis and are removed from the sample before analysis. Methods that utilize postcolumn derivatization of the amino acids are generally not affected by buffer components to the extent seen with precolumn

129 derivatization methods. It is desirable to limit the number of sample  
130 manipulations to reduce potential background contamination, to improve analyte  
131 recovery, and to reduce labor. Common techniques used to remove buffer  
132 components from protein samples include the following methods: (1) injecting  
133 the protein sample onto a reversed-phase HPLC system, removing the protein  
134 with a volatile solvent containing a sufficient organic component, and drying the  
135 sample in a vacuum centrifuge; (2) dialysis against a volatile buffer or water; (3)  
136 centrifugal ultrafiltration for buffer replacement with a volatile buffer or water;  
137 (4) precipitating the protein from the buffer using an organic solvent (e.g.,  
138 acetone); and (5) gel filtration.

139  
140

### **Internal Standards**

141  
142 It is recommended that an internal standard be used to monitor physical  
143 and chemical losses and variations during amino acid analysis. An accurately  
144 known amount of internal standard can be added to a protein solution prior to  
145 hydrolysis. The recovery of the internal standard gives the general recovery of  
146 the amino acids of the protein solution. Free amino acids, however, do not  
147 behave in the same way as protein-bound amino acids during hydrolysis because  
148 their rates of release or destruction are variable. Therefore, the use of an internal  
149 standard to correct for losses during hydrolysis may give unreliable results. It will  
150 be necessary to take this point under consideration when interpreting the results.  
151 Internal standards can also be added to the mixture of amino acids after hydrolysis  
152 to correct for differences in sample application and changes in reagent stability  
153 and flow rates. Ideally, an internal standard is an unnaturally occurring primary  
154 amino acid that is commercially available and inexpensive. It should also be  
155 stable during hydrolysis, its response factor should be linear with concentration,  
156 and it needs to elute with a unique retention time without overlapping other amino  
157 acids. Commonly used amino acid standards include norleucine, nitrotyrosine,  
158 and  $\alpha$ -aminobutyric acid.  
159

## 160 **Protein Hydrolysis**

161  
162 Hydrolysis of protein and peptide samples is necessary for amino acid  
163 analysis of these molecules. The glassware used for hydrolysis must be very  
164 clean to avoid erroneous results. Glove powders and fingerprints on hydrolysis  
165 tubes may cause contamination. To clean glass hydrolysis tubes, boil tubes for 1  
166 hour in 1 N hydrochloric acid or soak tubes in concentrated nitric acid or in a  
167 mixture of concentrated hydrochloric acid and concentrated nitric acid (1:1).  
168 Clean hydrolysis tubes are rinsed with high-purity water followed by a rinse with  
169 HPLC grade methanol, dried overnight in an oven, and stored covered until use.  
170 Alternatively, pyrolysis of clean glassware at 500°C for 4 hours may also be used  
171 to eliminate contamination from hydrolysis tubes. Adequate disposable  
172 laboratory material can also be used.

173 Acid hydrolysis is the most common method for hydrolyzing a protein  
174 sample before amino acid analysis. The acid hydrolysis technique can contribute  
175 to the variation of the analysis due to complete or partial destruction of several  
176 amino acids. Tryptophan is destroyed; serine and threonine are partially  
177 destroyed; methionine might undergo oxidation; and cysteine is typically  
178 recovered as cystine (but cystine recovery is usually poor because of partial  
179 destruction or reduction to cysteine). Application of adequate vacuum ( $\leq$  less  
180 than 200  $\mu\text{m}$  of mercury or 26.7 Pa) or introduction of an inert gas (argon) in the  
181 headspace of the reaction vessel can reduce the level of oxidative destruction. In  
182 peptide bonds involving isoleucine and valine the amido bonds of Ile-Ile, Val-Val,  
183 Ile-Val, and Val-Ile are partially cleaved; and asparagine and glutamine are  
184 deamidated, resulting in aspartic acid and glutamic acid, respectively. The loss of  
185 tryptophan, asparagine, and glutamine during an acid hydrolysis limits  
186 quantitation to 17 amino acids. Some of the hydrolysis techniques described are  
187 used to address these concerns. Some of the hydrolysis techniques described (i.e.,  
188 *Methods 4-11*) may cause modifications to other amino acids. Therefore, the  
189 benefits of using a given hydrolysis technique are weighed against the concerns  
190 with the technique and are tested adequately before employing a method other  
191 than acid hydrolysis.

192 A time-course study (i.e., amino acid analysis at acid hydrolysis times of  
193 24, 48, and 72 hours) is often employed to analyze the starting concentration of  
194 amino acids that are partially destroyed or slow to cleave. By plotting the  
195 observed concentration of labile amino acids (i.e., serine and threonine) versus  
196 hydrolysis time, the line can be extrapolated to the origin to determine the starting  
197 concentration of these amino acids. Time-course hydrolysis studies are also used  
198 with amino acids that are slow to cleave (e.g., isoleucine and valine). During the  
199 hydrolysis time course, the analyst will observe a plateau in these residues. The  
200 level of this plateau is taken as the residue concentration. If the hydrolysis time is  
201 too long, the residue concentration of the sample will begin to decrease, indicating  
202 destruction by the hydrolysis conditions.

203 An acceptable alternative to the time-course study is to subject an amino  
204 acid calibration standard to the same hydrolysis conditions as the test sample.  
205 The amino acid in free form may not completely represent the rate of destruction  
206 of labile amino acids within a peptide or protein during the hydrolysis. This is  
207 especially true for peptide bonds that are slow to cleave (e.g., Ile-Val bonds).  
208 However, this technique will allow the analyst to account for some residue  
209 destruction. Microwave acid hydrolysis has been used and is rapid but requires  
210 special equipment as well as special precautions. The optimal conditions for  
211 microwave hydrolysis must be investigated for each individual protein/peptide  
212 sample. The microwave hydrolysis technique typically requires only a few  
213 minutes, but even a deviation of one minute may give inadequate results (e.g.,  
214 incomplete hydrolysis or destruction of labile amino acids). Complete  
215 proteolysis, using a mixture of proteases, has been used but can be complicated,  
216 requires the proper controls, and is typically more applicable to peptides than  
217 proteins.

218 NOTE—During initial analyses of an unknown protein, experiments with  
219 various hydrolysis time and temperature conditions are conducted to determine  
220 the optimal conditions.

221  
222  
223

## METHOD 1

224 Acid hydrolysis using hydrochloric acid containing phenol is the most  
225 common procedure used for protein/peptide hydrolysis preceding amino acid  
226 analysis. The addition of phenol to the reaction prevents the halogenation of  
227 tyrosine.

228 **Hydrolysis Solution:** 6 N hydrochloric acid containing 0.1% to 1.0% of  
229 phenol.

230 **Procedure—**

231 *Liquid Phase Hydrolysis*—Place the protein or peptide sample in a  
232 hydrolysis tube, and dry. [NOTE—The sample is dried so that water in the sample  
233 will not dilute the acid used for the hydrolysis.] Add 200  $\mu\text{L}$  of *Hydrolysis*  
234 *Solution* per 500  $\mu\text{g}$  of lyophilized protein. Freeze the sample tube in a dry ice-  
235 acetone bath, and flame seal in vacuum. Samples are typically hydrolyzed at  
236 110°C for 24 hours in vacuum or inert atmosphere to prevent oxidation. Longer  
237 hydrolysis times (e.g., 48 and 72 hours) are investigated if there is a concern that  
238 the protein is not completely hydrolyzed.

239 *Vapor Phase Hydrolysis*—This is one of the most common acid hydrolysis  
240 procedures, and it is preferred for microanalysis when only small amounts of the  
241 sample are available. Contamination of the sample from the acid reagent is also  
242 minimized by using vapor phase hydrolysis. Place vials containing the dried  
243 samples in a vessel that contains an appropriate amount of *Hydrolysis Solution*.  
244 The *Hydrolysis Solution* does not come in contact with the test sample. Apply an  
245 inert atmosphere or vacuum ( $\leq$  less than 200  $\mu\text{m}$  of mercury or 26.7 Pa) to the  
246 headspace of the vessel, and heat to about 110°C for a 24-hour hydrolysis time.  
247 Acid vapor hydrolyzes the dried sample. Any condensation of the acid in the  
248 sample vials is minimized. After hydrolysis, dry the test sample in vacuum to  
249 remove any residual acid.

250 METHOD 2

251  
252 Tryptophan oxidation during hydrolysis is decreased by using  
253 mercaptoethanesulfonic acid (MESA) as the reducing acid.

254 **Hydrolysis Solution:** 2.5 M MESA solution.

255           **Vapor Phase Hydrolysis**—About 1 to 100 µg of the protein/peptide  
256 under test is dried in a hydrolysis tube. The hydrolysis tube is placed in a larger  
257 tube with about 200 µL of the *Hydrolysis Solution*. The larger tube is sealed in  
258 vacuum (about 50 µm of mercury or 6.7 Pa) to vaporize the *Hydrolysis Solution*.  
259 The hydrolysis tube is heated to 170°C to 185°C for about 12.5 minutes. After  
260 hydrolysis, the hydrolysis tube is dried in vacuum for 15 minutes to remove the  
261 residual acid.

262  
263

### METHOD 3

264  
265

Tryptophan oxidation during hydrolysis is prevented by using thioglycolic acid (TGA) as the reducing acid.

267           **Hydrolysis Solution**—A solution containing 7 M hydrochloric acid, 10%  
268 of trifluoroacetic acid, 20% of thioglycolic acid, and 1% of phenol.

269           **Vapor Phase Hydrolysis**—About 10 to 50 µg of the protein/peptide  
270 under test is dried in a sample tube. The sample tube is placed in a larger tube  
271 with about 200 µL of the *Hydrolysis Solution*. The larger tube is sealed in  
272 vacuum (about 50 µm of mercury or 6.7 Pa) to vaporize the TGA. The sample  
273 tube is heated to 166°C for about 15 to 30 minutes. After hydrolysis, the sample  
274 tube is dried in vacuum for 5 minutes to remove the residual acid. Recovery of  
275 tryptophan by this method may be dependent on the amount of sample present.

276  
277

### METHOD 4

278  
279

Cysteine-cystine and methionine oxidation is performed with performic acid before the protein hydrolysis.

281           **Oxidation Solution**—The performic acid is prepared fresh by mixing  
282 formic acid and 30 percent hydrogen peroxide (9:1), and incubated at room  
283 temperature for 1 hour.

284           **Procedure**—The protein/peptide sample is dissolved in 20 µL of formic  
285 acid, and heated at 50°C for 5 minutes; then 100 µL of the *Oxidation Solution* is  
286 added. The oxidation is allowed to proceed for 10 to 30 minutes. In this reaction,  
287 cysteine is converted to cysteic acid and methionine is converted to methionine

288 sulfone. The excess reagent is removed from the sample in a vacuum centrifuge.  
289 This technique may cause modifications to tyrosine residues in the presence of  
290 halides. The oxidized protein can then be acid hydrolyzed using *Method 1* or  
291 *Method 2*.

292  
293

#### METHOD 5

294  
295

Cysteine-cystine oxidation is accomplished during the liquid phase  
296 hydrolysis with sodium azide.

297 **Hydrolysis Solution:** 6 N hydrochloric acid containing 0.2% of phenol,  
298 to which is added sodium azide to obtain a final concentration of 0.2% (w/v). The  
299 added phenol prevents halogenation of tyrosine.

300 **Liquid Phase Hydrolysis**—The protein/peptide hydrolysis is conducted  
301 at about 110°C for 24 hours. During the hydrolysis, the cysteine-cystine present  
302 in the sample is converted to cysteic acid by the sodium azide present in the  
303 *Hydrolysis Solution*. This technique allows better tyrosine recovery than *Method*  
304 *4*, but it is not quantitative for methionine. Methionine is converted to a mixture  
305 of the parent methionine and its two oxidative products, methionine sulfoxide and  
306 methionine sulfone.

307  
308

#### METHOD 6

309  
310

Cysteine-cystine oxidation is accomplished with dimethyl sulfoxide  
311 (DMSO).

312 **Hydrolysis Solution:** 6 N hydrochloric acid containing 0.1% to 1.0% of  
313 phenol, to which DMSO is added to obtain a final concentration of 2% (v/v).

314 **Vapor Phase Hydrolysis**—The protein/peptide hydrolysis is conducted at  
315 about 110°C for 24 hours. During the hydrolysis, the cysteine-cystine present in  
316 the sample is converted to cysteic acid by the DMSO present in the *Hydrolysis*  
317 *Solution*. As an approach to limit variability and compensate for partial  
318 destruction, it is recommended to evaluate the cysteic acid recovery from  
319 oxidative hydrolyses of standard proteins containing 1 to 8 mol of cysteine per  
320 mol protein. The response factors from protein/peptide hydrolysates are typically

321 about 30% lower than those for nonhydrolyzed cysteic acid standards. Because  
322 histidine, methionine, tyrosine, and tryptophan are also modified, a complete  
323 compositional analysis is not obtained with this technique.

324

325

## METHOD 7

326

327 Cysteine-cystine reduction and alkylation is accomplished by a vapor  
328 phase pyridylethylation reaction.

329

330 **Reducing Solution**—Transfer 83.3  $\mu\text{L}$  of pyridine, 16.7  $\mu\text{L}$  of 4-  
331 vinylpyridine, 16.7  $\mu\text{L}$  of tributylphosphine, and 83.3  $\mu\text{L}$  of water to a suitable  
332 container, and mix.

333

334 **Procedure**—Add the protein/peptide (between 1 and 100  $\mu\text{g}$ ) to a  
335 hydrolysis tube, and place in a larger tube. Transfer the *Reducing Solution* to the  
336 large tube, seal in vacuum (about 50  $\mu\text{m}$  of mercury or 6.7 Pa), and incubate at  
337 about 100°C for 5 minutes. Then remove the inner hydrolysis tube, and dry it in a  
338 vacuum desiccator for 15 minutes to remove residual reagents. The  
339 pyridylethylated protein/peptide can then be acid hydrolyzed using previously  
340 described procedures. The pyridylethylation reaction is performed  
341 simultaneously with a protein standard sample containing 1 to 8 mol of cysteine  
342 per mol protein to improve accuracy in the pyridylethyl-cysteine recovery.  
343 Longer incubation times for the pyridylethylation reaction can cause  
344 modifications to the  $\alpha$ -amino terminal group and the  $\epsilon$ -amino group of lysine in  
345 the protein.

346

## METHOD 8

347

348 Cysteine-cystine reduction and alkylation is accomplished by a liquid  
349 phase pyridylethylation reaction.

350

351 **Stock Solutions**—Prepare and filter three solutions: 1 M Tris  
hydrochloride (pH 8.5) containing 4 mM edetate disodium (*Stock Solution A*), 8  
M guanidine hydrochloride (*Stock Solution B*), and 10% of 2-mercaptoethanol in  
water (*Stock Solution C*).

352           **Reducing Solution**—Prepare a mixture of *Stock Solution B* and *Stock*  
353 *Solution A* (3:1) to obtain a buffered solution of 6 M guanidine hydrochloride in  
354 0.25 M Tris hydrochloride.

355           **Procedure**—Dissolve about 10 µg of the test sample in 50 µL of the  
356 *Reducing Solution*, and add about 2.5 µL of *Stock Solution C*. Store under  
357 nitrogen or argon for 2 hours at room temperature in the dark. To achieve the  
358 pyridylethylation reaction, add about 2 µL of 4-vinylpyridine to the protein  
359 solution, and incubate for an additional 2 hours at room temperature in the dark.  
360 The protein/peptide is desalted by collecting the protein/peptide fraction from a  
361 reversed-phase HPLC separation. The collected sample can be dried in a vacuum  
362 centrifuge before acid hydrolysis.

363  
364

#### METHOD 9

365  
366           Cysteine-cystine reduction and alkylation is accomplished by a liquid  
367 phase carboxymethylation reaction.

368           **Stock Solutions**—Prepare as directed for *Method 8*.

369           **Carboxymethylation Solution**—Prepare a solution containing 100 mg of  
370 iodoacetamide per mL of alcohol.

371           **Buffer Solution**—Use the *Reducing Solution*, prepared as directed for  
372 *Method 8*.

373           **Procedure**—Dissolve the test sample in 50 µL of the *Buffer Solution*, and  
374 add about 2.5 µL of *Stock Solution C*. Store under nitrogen or argon for 2 hours  
375 at room temperature in the dark. Add the *Carboxymethylation Solution* in a ratio  
376 1.5 fold per total theoretical content of thiols, and incubate for an additional 30  
377 minutes at room temperature in the dark. [NOTE—If the thiol content of the  
378 protein is unknown, then add 5 µL of 100 mM iodoacetamide for every 20 nmol  
379 of protein present.] The reaction is stopped by adding excess of 2-  
380 mercaptoethanol. The protein/peptide is desalted by collecting the protein/peptide  
381 fraction from a reversed-phase HPLC separation. The collected sample can be  
382 dried in a vacuum centrifuge before acid hydrolysis. The *S*-carboxyamidomethyl-

383 cysteine formed will be converted to *S*-carboxymethylcysteine during acid  
384 hydrolysis.

385 METHOD 10

386  
387 Cysteine-cystine is reacted with dithiodiglycolic acid or dithiodipropionic  
388 acid to produce a mixed disulfide. [NOTE—The choice of dithiodiglycolic acid or  
389 dithiodipropionic acid depends on the required resolution of the amino acid  
390 analysis method.]

391 **Reducing Solution**—A solution containing 10 mg of dithiodiglycolic acid  
392 (or dithiodipropionic acid) per mL of 0.2 M sodium hydroxide.

393 **Procedure**—Transfer about 20 µg of the test sample to a hydrolysis tube,  
394 and add 5 µL of the *Reducing Solution*. Add 10 µL of isopropyl alcohol, and then  
395 remove all of the sample liquid by vacuum centrifugation. The sample is then  
396 hydrolyzed using *Method 1*. This method has the advantage that other amino acid  
397 residues are not derivatized by side reactions, and the sample does not need to be  
398 desalted prior to hydrolysis.

399  
400

401 METHOD 11

402 Asparagine and glutamine are converted to aspartic acid and glutamic  
403 acid, respectively, during acid hydrolysis. Asparagine and aspartic acid residues  
404 are added and represented by *Asx*, while glutamine and glutamic acid residues are  
405 added and represented by *Glx*. Proteins/peptides can be reacted with bis(1,1-  
406 trifluoroacetoxy)iodobenzene (BTI) to convert the asparagine and glutamine  
407 residues to diaminopropionic acid and diaminobutyric acid residues, respectively,  
408 upon acid hydrolysis. These conversions allow the analyst to determine the  
409 asparagine and glutamine content of a protein/peptide in the presence of aspartic  
410 acid and glutamic acid residues.

411 **Reducing Solutions**—Prepare and filter three solutions: a solution of 10  
412 mM trifluoroacetic acid (*Solution A*), a solution of 5 M guanidine hydrochloride  
413 and 10 mM trifluoroacetic acid (*Solution B*), and a freshly prepared solution of  
414 dimethylformamide containing 36 mg of BTI per mL (*Solution C*).

415           **Procedure**—In a clean hydrolysis tube, transfer about 200 µg of the test  
416 sample, and add 2 mL of *Solution A* or *Solution B* and 2 mL of *Solution C*. Seal  
417 the hydrolysis tube in vacuum. Heat the sample at 60°C for 4 hours in the dark.  
418 The sample is then dialyzed with water to remove the excess reagents. Extract the  
419 dialyzed sample three times with equal volumes of n-butyl acetate, and then  
420 lyophilize. The protein can then be acid hydrolyzed using previously described  
421 procedures. The  $\alpha,\beta$ -diaminopropionic and  $\alpha,\gamma$ -diaminobutyric acid residues do  
422 not typically resolve from the lysine residues upon ion-exchange chromatography  
423 based on amino acid analysis. Therefore, when using ion-exchange as the mode  
424 of amino acid separation, the asparagine and glutamine contents are the  
425 quantitative difference in the aspartic acid and glutamic acid content assayed with  
426 underivatized and BTI-derivatized acid hydrolysis. [NOTE—The threonine,  
427 methionine, cysteine, tyrosine, and histidine assayed content can be altered by  
428 BTI derivatization; a hydrolysis without BTI will have to be performed if the  
429 analyst is interested in the composition of these other amino acid residues of the  
430 protein/peptide.]

431  
432

### 433           **Methodologies of Amino Acid Analysis General Principles**

434           Many amino acid analysis techniques exist, and the choice of any one  
435 technique often depends on the sensitivity required from the assay. In general,  
436 about one-half of the amino acid analysis techniques employed rely on the  
437 separation of the free amino acids by ion-exchange chromatography followed by  
438 postcolumn derivatization (e.g., with ninhydrin or *o*-phthalaldehyde). Postcolumn  
439 detection techniques can be used with samples that contain small amounts of  
440 buffer components, such as salts and urea, and generally require between 5 and 10  
441 µg of protein sample per analysis. The remaining amino acid techniques typically  
442 involve precolumn derivatization of the free amino acids (e.g., phenyl  
443 isothiocyanate; 6-amino-quinolyl-N-hydroxysuccinimidyl carbanate or *o*-  
444 phthalaldehyde; (dimethylamino) azobenzenesulfonyl chloride; 9-  
445 fluorenylmethylchloroformate; and, 7-fluoro-4-nitrobenzo-2-oxa-1,3-diazole)  
446 followed by reversed-phase HPLC. Precolumn derivatization techniques are very

447 sensitive and usually require between 0.5 and 1.0  $\mu\text{g}$  of protein sample per  
448 analysis but may be influenced by buffer salts in the samples. Precolumn  
449 derivatization techniques may also result in multiple derivatives of a given amino  
450 acid, which complicates the result interpretation. Postcolumn derivatization  
451 techniques are generally influenced less by performance variation of the assay  
452 than precolumn derivatization techniques.

453       The following *Methods* may be used for quantitative amino acid analysis.  
454 Instruments and reagents for these procedures are available commercially.  
455 Furthermore, many modifications of these methodologies exist with different  
456 reagent preparations, reaction procedures, chromatographic systems, etc. Specific  
457 parameters may vary according to the exact equipment and procedure used.  
458 Many laboratories will utilize more than one amino acid analysis technique to  
459 exploit the advantages offered by each. In each of these *Methods*, the analog  
460 signal is visualized by means of a data acquisition system, and the peak areas are  
461 integrated for quantification purposes.

462  
463               METHOD 1 – Postcolumn Ninhydrin Detection General Principle

464  
465       Ion-exchange chromatography with postcolumn ninhydrin detection is one  
466 of the most common methods employed for quantitative amino acid analysis. As  
467 a rule, a Li-based cation-exchange system is employed for the analysis of the  
468 more complex physiological samples, and the faster Na-based cation-exchange  
469 system is used for the more simplistic amino acid mixtures obtained with protein  
470 hydrolysates (typically containing 17 amino acid components). Separation of the  
471 amino acids on an ion-exchange column is accomplished through a combination  
472 of changes in pH and cation strength. A temperature gradient is often employed  
473 to enhance separation.

474       When the amino acid reacts with ninhydrin, the reactant has characteristic  
475 purple or yellow color. Amino acids, except imino acid, give a purple color, and  
476 show the maximum absorption at 570 nm. The imino acids such as proline give a  
477 yellow color, and show the maximum absorption at 440 nm. The postcolumn  
478 reaction between ninhydrin and amino acid eluted from column is monitored at

479 440 and 570 nm, and the chromatogram obtained is used for the determination of  
480 amino acid composition.

481 Detection limit is considered to be 10 pmol for most of the amino acid  
482 derivatives, but 50 pmol for proline. Response linearity is obtained in the range  
483 of 20 to 500 pmol with correlation coefficients exceeding 0.999. To obtain good  
484 composition data, samples larger than 1 µg before hydrolysis are best suited for  
485 this amino acid analysis of protein/peptide.

486

487 METHOD 2 – Postcolumn OPA Fluorometric Detection General Principle

488

489 *o*-Phthalaldehyde (OPA) reacts with primary amines in the presence of  
490 thiol compound, to form highly fluorescent isoindole products. This reaction is  
491 utilized for the postcolumn derivatization in analysis of amino acids by ion-  
492 exchange chromatography. The rule of the separation is the same as *Method 1*.  
493 Instruments and reagents for this form of amino acid analysis are available  
494 commercially. Many modifications of this methodology exist.

495 Although OPA does not react with secondary amines (imino acids such as  
496 proline) to form fluorescent substances, the oxidation with sodium hypochlorite  
497 allows secondary amines to react with OPA. The procedure employs a strongly  
498 acidic cation-exchange column for separation of free amino acids followed by  
499 postcolumn oxidation with sodium hypochlorite and postcolumn derivatization  
500 using OPA and thiol compound such as *N*-acetyl-L-cysteine and 2-  
501 mercaptoethanol. The derivatization of primary amino acids are not noticeably  
502 affected by the continuous supply of sodium hypochlorite.

503 Separation of the amino acids on an ion-exchange column is accomplished  
504 through a combination of changes in pH and cation strength. After postcolumn  
505 derivatization of eluted amino acids with OPA, the reactant passes through the  
506 fluorometric detector. Fluorescence intensity of OPA-derivatized amino acids are  
507 monitored with an excitation wavelength of 348 nm and an emission wavelength  
508 of 450 nm.

509 Detection limit is considered to be a few tens of picomole level for most of  
510 the amino acid derivatives. Response linearity is obtained in the range of a few

511 picomole level to a few tens of nanomole level. To obtain good compositional  
512 data, the starting with greater than 500 ng of sample before hydrolysis is best  
513 suited for the amino acid analysis of protein/peptide.

514

515           METHOD 3 – Precolumn PITC Derivatization General Principle

516

517           Phenylisothiocyanate (PITC) reacts with amino acids to form  
518 phenylthiocarbamyl (PTC) derivatives which can be detected with high sensitivity  
519 at 245 nm. Therefore, precolumn derivatization of amino acids with PITC  
520 followed by a reversed-phase HPLC separation with UV detection is used to  
521 analyze the amino acid composition.

522           After the reagent is removed under vacuum, the derivatized amino acids  
523 can be stored dry and frozen for several weeks with no significant degradation. If  
524 the solution for injection is kept cold, no noticeable loss in chromatographic  
525 response occurs after three days.

526           Separation of the PTC-amino acids on a reversed-phase HPLC with ODS  
527 column is accomplished through a combination of changes in concentrations of  
528 acetonitrile and buffer ionic strength. PTC-amino acids eluted from column are  
529 monitored at 254 nm.

530           Detection limit is considered to be 1 pmol for most of the amino acid  
531 derivatives. Response linearity is obtained in the range of 20 to 500 pmol with  
532 correlation coefficients exceeding 0.999. To obtain good compositional data,  
533 samples larger than 500 ng of protein/peptide before hydrolysis is best suited for  
534 this amino acid analysis of proteins/peptides.

535

536           METHOD 4 – Precolumn AQC Derivatization General Principle

537

538           Precolumn derivatization of amino acids with 6-aminoquinolyl-*N*-  
539 hydroxysuccinimidyl carbamate (AQC) followed by reversed-phase HPLC  
540 separation with fluorometric detection is used.

541           6-Aminoquinolyl-*N*-hydroxysuccinimidyl carbamate (AQC) reacts with  
542 amino acids to form stable, fluorescent unsymmetric urea derivatives (AQC-  
543 amino acids) which are readily amenable to analysis by reversed-phase HPLC.

544 Therefore, precolumn derivatization of amino acids with AQC followed by  
545 reversed-phase HPLC separation is used to analyze the amino acid composition.

546 Separation of the AQC-amino acids on ODS column is accomplished  
547 through a combination of changes in concentrations of acetonitrile and salt.  
548 Selective fluorescence detection of the derivatives with excitation wavelength at  
549 250 nm and emission wavelength at 395 nm allows for the direct injection of the  
550 reaction mixture with no significant interference from the only major fluorescent  
551 reagent by-product, 6-aminoquinoline. Excess reagent is rapidly hydrolyzed ( $t_{1/2}$   
552 <15 seconds) to yield 6-aminoquinoline, *N*-hydroxysuccinimide and carbon  
553 dioxide, and after 1 minute no further derivatization can take place.

554 Peak areas for AQC-amino acids are essentially unchanged for at least 1  
555 week at room temperature, and the derivatives have more than sufficient stability  
556 to allow for overnight automated chromatographic analysis.

557 Detection limit is considered to be ranging from *ca.* 40 to 320 fmol for  
558 each amino acid, except for Cys. Detection limit for Cys is approximately 800  
559 fmol. Response linearity is obtained in the range of 2.5 to 200  $\mu$ M with  
560 correlation coefficients exceeding 0.999. Good compositional data could be  
561 obtained from the analysis of derivatized protein hydrolysates containing as little  
562 as 30 ng of protein/peptide.

563

#### 564 METHOD 5 – Precolumn OPA Derivatization General Principle

565

566 Precolumn derivatization of amino acids with *o*-phthalaldehyde (OPA)  
567 followed by reversed-phase HPLC separation with fluorometric detection is used.  
568 This technique does not detect amino acids that exist as secondary amines (e.g.,  
569 proline).

570 *o*-Phthalaldehyde (OPA) in conjunction with a thiol reagent reacts with  
571 primary amine groups to form highly fluorescent isoindole products. 2-  
572 Mercaptoethanol or 3-mercaptopropionic acid can be used as the thiol. OPA itself  
573 does not fluoresce and consequently produces no interfering peaks. In addition,  
574 its solubility and stability in aqueous solution, along with the rapid kinetics for the  
575 reaction, make it amenable to automated derivatization and analysis using an

576 autosampler to mix the sample with the reagent. However, lack of reactivity with  
577 secondary amino acids has been predominant drawback. This method does not  
578 detect amino acids that exist as secondary amines (e.g., proline). To compensate  
579 for this drawback, this technique may be combined with another technique  
580 described in *Method 7* or *Method 8*.

581         Precolumn derivatization of amino acids with OPA is followed by a  
582 reversed-phase HPLC separation. Because of the instability of the OPA-amino  
583 acid derivative, HPLC separation and analysis are performed immediately  
584 following derivatization. The liquid chromatograph is equipped with a  
585 fluorometric detector for the detection of derivatized amino acids. Fluorescence  
586 intensity of OPA-derivatized amino acids is monitored with an excitation  
587 wavelength of 348 nm and an emission wavelength of 450 nm.

588         Detection limits as low as 50 fmol via fluorescence have been reported,  
589 although the practical limit of analysis remains at 1 pmol.

590

#### 591         METHOD 6 – Precolumn DABS-Cl Derivatization General 592         Principle

592

593         Precolumn derivatization of amino acids with  
594 (dimethylamino)azobenzenesulfonyl chloride (DABS-Cl) followed by reversed-  
595 phase HPLC separation with visible light detection is used.

596         (Dimethylamino)azobenzenesulfonyl chloride (DABS-Cl) is a  
597 chromophoric reagent employed for the labeling of amino acids. Amino acids  
598 labeled with DABS-Cl (DABS-amino acids) are highly stable and show the  
599 maximum absorption at 436 nm.

600         DABS-amino acids, all 19 naturally occurring amino acids derivatives,  
601 can be separated on an ODS column of a reversed-phase HPLC by employing  
602 gradient systems consisting of acetonitrile and aqueous buffer mixture. Separated  
603 DABS-amino acids eluted from column are detected at 436 nm in the visible  
604 region.

605         This *Method* can analyze the imino acids such as proline together with the  
606 amino acids at the same degree of sensitivity, DABS-Cl derivatization method  
607 permits the simultaneous quantification of tryptophan residues by previous

608 hydrolysis of the protein/peptide with sulfonic acids such as  
609 mercaptoethanesulfonic acid, *p*-toluenesulfonic acid or methanesulfonic acid  
610 described under *Method 2* in “Protein Hydrolysis”. The other acid-labile residues,  
611 asparagine and glutamine, can also be analysed by previous conversion into  
612 diaminopropionic acid and diaminobutyric acid, respectively, by treatment of  
613 protein/peptide with BTI described under *Method 11* in “Protein Hydrolysis”.

614 The non-proteinogenic amino acid, norleucine cannot be used as internal  
615 standard in this method, as this compound is eluted in a chromatographic region  
616 crowded with peaks of primary amino acids. Nitrotyrosine can be used as an  
617 internal standard, because it is eluted in a clean region.

618 Detection limit of DABS-amino acid is about 1 pmol. As little as 2 to 5  
619 pmol of an individual DABS-amino acid can be quantitatively analysed with  
620 reliability, and only 10 to 30 ng of the dabsylated protein hydrolysate is required  
621 for each analysis.

622

#### 623 METHOD 7 – Precolumn FMOC-Cl Derivatization General 624 Principle

624

625 Precolumn derivatization of amino acids with 9-fluorenylmethyl  
626 chloroformate (FMOC-Cl) followed by reversed-phase HPLC separation with  
627 fluorometric detection is used.

628 9-Fluorenylmethyl chloroformate (FMOC-Cl) reacts with both primary  
629 and secondary amino acids to form highly fluorescent products. The reaction of  
630 FMOC-Cl with amino acid proceeds under mild conditions in aqueous solution  
631 and is completed in 30 seconds. The derivatives are stable, only the histidine  
632 derivative showing any breakdown. Although FMOC-Cl is fluorescent itself, the  
633 reagent excess and fluorescent side-products can be eliminated without loss of  
634 FMOC-amino acids.

635 FMOC-amino acids are separated by a reversed-phase HPLC using ODS  
636 column. The separation is carried out by gradient elution varied linearly from a  
637 mixture of acetonitrile methanol and acetic acid buffer (10:40:50) to a mixture of  
638 acetonitrile and acetic acid buffer (50:50), and 20 amino acid derivatives are  
639 separated in 20 minutes. Each derivative eluted from column is monitored by a

640 fluorometric detector set at an excitation wavelength of 260 nm and an emission  
641 wavelength of 313 nm.

642 The detection limit is in the low fmol range. A linearity range of 0.1 to 50  
643  $\mu\text{M}$  is obtained for most of the amino acids.

644  
645 METHOD 8 – Precolumn NBD-F Derivatization General Principle

646  
647 Precolumn derivatization of amino acids with 7-fluoro-4-nitrobenzo-2-  
648 oxa-1,3-diazole (NBD-F) followed by reversed-phase HPLC separation with  
649 fluorometric detection is used.

650 7-fluoro-4-nitrobenzo-2-oxa-1,3-diazole (NBD-F) reacts with both  
651 primary and secondary amino acids to form highly fluorescent products. Amino  
652 acids are derivatized with NBD-F by heating to 60°C for 5 minutes.

653 NBD-amino acid derivatives are separated on an ODS column of a  
654 reversed-phase HPLC by employing gradient elution system consisting of  
655 acetonitrile and aqueous buffer mixture, and 17 amino acid derivatives are  
656 separated in 35 minutes.  $\epsilon$ -Aminocaproic acid can be used as an internal standard,  
657 because it is eluted in a clean chromatographic region. Each derivative eluted  
658 from column is monitored by a fluorometric detector set at an excitation  
659 wavelength of 480 nm and an emission wavelength of 530 nm.

660 The sensitivity of this method is almost the same as for precolumn OPA  
661 derivatization method (*Method 5*), excluding proline to which OPA is not  
662 reactive, and might be advantageous for NBD-F against OPA. The detection limit  
663 for each amino acid is about 10 fmol. Profile analysis can be achieved for about  
664 1.5  $\mu\text{g}$  of protein hydrolysates in the final precolumn labeling reaction mixture for  
665 HPLC.

## 666 667 **Data Calculation and Analysis**

668  
669 When determining the amino acid content of a protein/peptide  
670 hydrolysate, it should be noted that the acid hydrolysis step destroys tryptophan  
671 and cysteine. Serine and threonine are partially destroyed by acid hydrolysis,  
672 while isoleucine and valine residues may be only partially cleaved. Methionine

673 can undergo oxidation during acid hydrolysis, and some amino acids (e.g., glycine  
674 and serine) are common contaminants. Application of adequate vacuum ( $\leq$  not  
675 less than 200  $\mu\text{m}$  of mercury or 26.7 Pa) or introduction of inert gas (argon) in the  
676 headspace of the reaction vessel during vapor phase hydrolysis can reduce the  
677 level of oxidative destruction. Therefore, the quantitative results obtained for  
678 cysteine, tryptophan, threonine, isoleucine, valine, methionine, glycine, and serine  
679 from a protein/peptide hydrolysate may be variable and may warrant further  
680 investigation and consideration.

681  
682

#### CALCULATIONS

683  
684

**Amino Acid Mole Percent**—This is the number of specific amino acid  
685 residues per 100 residues in a protein. This result may be useful for evaluating  
686 amino acid analysis data when the molecular weight of the protein under  
687 investigation is unknown. This information can be used to corroborate the  
688 identity of a protein/peptide and has other applications. Carefully identify and  
689 integrate the peaks obtained as directed for each *Procedure*. Calculate the mole  
690 percent for each amino acid present in the test sample by the formula:

691

$$100r_U/r,$$

692 in which  $r_U$  is the peak response, in nmol, of the amino acid under test; and  $r$  is  
693 the sum of peak responses, in nmol, for all amino acids present in the test sample.  
694 Comparison of the mole percent of the amino acids under test to data from known  
695 proteins can help establish or corroborate the identity of the sample protein.

696 **Unknown Protein Samples**—This data analysis technique can be used to  
697 estimate the protein concentration of an unknown protein sample using the amino  
698 acid analysis data. Calculate the mass, in  $\mu\text{g}$ , of each recovered amino acid by the  
699 formula:

700

$$mM_W/1000,$$

701 in which  $m$  is the recovered quantity, in nmol, of the amino acid under test; and  
702  $M_W$  is the average molecular weight for that amino acid, corrected for the weight  
703 of the water molecule that was eliminated during peptide bond formation. The  
704 sum of the masses of the recovered amino acids will give an estimate of the total

705 mass of the protein analyzed after appropriate correction for partially and  
706 completely destroyed amino acids. If the molecular weight of the unknown  
707 protein is available (i.e., by SDS-PAGE analysis or mass spectroscopy), the  
708 amino acid composition of the unknown protein can be predicted. Calculate the  
709 number of residues of each amino acid by the formula:

$$710 \quad m/(1000M/M_{WT}),$$

711 in which  $m$  is the recovered quantity, in nmol, of the amino acid under test;  $M$  is  
712 the total mass, in  $\mu\text{g}$ , of the protein; and  $M_{WT}$  is the molecular weight of the  
713 unknown protein.

714 **Known Protein Samples**—This data analysis technique can be used to  
715 investigate the amino acid composition and protein concentration of a protein  
716 sample of known molecular weight and amino acid composition using the amino  
717 acid analysis data. When the composition of the protein being analyzed is known,  
718 one can exploit the fact that some amino acids are recovered well, while other  
719 amino acid recoveries may be compromised because of complete or partial  
720 destruction (e.g., tryptophan, cysteine, threonine, serine, methionine), incomplete  
721 bond cleavage (i.e., for isoleucine and valine) and free amino acid contamination  
722 (i.e., by glycine and serine).

723 Because those amino acids that are recovered best represent the protein,  
724 these amino acids are chosen to quantify the amount of protein. Well-recovered  
725 amino acids are, typically, aspartate-asparagine, glutamate-glutamine, alanine,  
726 leucine, phenylalanine, lysine, and arginine. This list can be modified based on  
727 experience with one's own analysis system. Divide the quantity, in nmol, of each  
728 of the well-recovered amino acids by the expected number of residues for that  
729 amino acid to obtain the protein content based on each well-recovered amino acid.  
730 Average the protein content results calculated. The protein content determined  
731 for each of the well-recovered amino acids should be evenly distributed about the  
732 mean. Discard protein content values for those amino acids that have an  
733 unacceptable deviation from the mean. Typically  $\geq$  greater than 5% variation  
734 from the mean is considered unacceptable. Recalculate the mean protein content  
735 from the remaining values to obtain the protein content of the sample. Divide the

736 content of each amino acid by the calculated mean protein content to determine  
737 the amino acid composition of the sample by analysis.

738 Calculate the relative compositional error, in percentage, by the formula:

739 
$$100m/m_s,$$

740 in which  $m$  is the experimentally determined quantity, in nmol per amino acid  
741 residue, of the amino acid under test; and  $m_s$  is the known residue value for that  
742 amino acid. The average relative compositional error is the average of the  
743 absolute values of the relative compositional errors of the individual amino acids,  
744 typically excluding tryptophan and cysteine from this calculation. The average  
745 relative compositional error can provide important information on the stability of  
746 analysis run over time. The agreement in the amino acid composition between the  
747 protein sample and the known composition can be used to corroborate the identity  
748 and purity of the protein in the sample.

749